JGL d.d.
Svilno 20
51000 Rijeka
OIB: 20950636972

Rijeka, 19. listopada 2021.

SVIM ZAINTERSIRANIM GOSPODARSKIM SUBJEKTIMA

Predmet: Odgovor na upit u postupku nabave s obveznom objavom za predmet nabave: Nabava opreme i uređaja za pilot pogon: Grupa 1 - Specijalizirana radna površina za rad s opasnim kemikalijama (digestori) i Grupa 2 – Mješači

Naručitelj je u predmetnom postupku nabave zaprimio upit zainteresiranog gospodarskog subjekta sljedećeg sadržaja:

UPIT 1:

„Grupa 1 – digestori,
zahtijevane karakteristike 5 i 11 – ne spominje se regulacija protoka zraka putem klapne, niti eventualni uvjet da bude od PPsEL-a.
Da li je to potrebno staviti u ponudu, ili ostati samo na indikaciji i alarmu? Razlika u cijeni je značajna!

Zahtijevana karakteristika 9 – komora za špricanje. Da li se navedena komora spaja na odsis digestora ili mora imati svoj ventilator?

ODGOVOR NARUČITELJA:

Digestori moraju biti opremljeni sa sustavom kontrole i održavanja količine zraka u optoku (regulacija protoka zraka putem klapne od PPsEL-a).

Komora za špricanje se spaja na odsis digestora.

UPIT 2:

Prema otvorenom natječaju „European structural and investment funds“ zainteresirani smo za nuđenje projekta (RINO SPREJ, broj ugovora KK.01.2.1.02.0111).

S obzirom na trenutnu situaciju na tržištu, nismo uspjeli pribaviti svu potrebnu dokumentaciju, odnosno sve potrebne ponude naših dobavljača opreme.

Molim Vas prema navedenom da nam dozvolite slanje ponude nakon navedenog roka za dostavu ponude.

Ponudu bi, naravno ukoliko je to sa vaše strane prihvatljivo poslali do 31.10.2021.

ODGOVOR NARUČITELJA:

Ponuda zaprimljena izvan roka za dostavu ponuda je nepravilna i neće se uzimati u obzir.

Međutim, nastavno na zahtjev zainteresiranog gospodarskog subjekta Naručitelj produljuje rok za dostavu ponuda do 02.11. 2021. godinu do 14:00 sati.

Prethodni odgovor i 1. Izmjenu Poziva Naručitelj javno objavljuje na istim internetskim stranicama na kojima je objavljena i osnovna dokumentacija te isti čini sastavni dio natječajne dokumentacije.

Naručitelj poziva sve zainteresirane gospodarske subjekte da pročitaju i prate sva pojašnjenja koja su objavljena u sklopu ovog postupka nabave.

Odbor za nabavu

JGL d.d.
Svilno 20
51000 Rijeka
Identification / Registration Number: 20950636972

Rijeka, October 19th 2021.

TO ALL INTERESTED BUSINESS ENTITIES

Subject: *Response to the request in the procurement procedure with mandatory publication for the subject of procurement: Procurement of equipment and devices for pilot plant. Lot 1 - Specialized work surface for working with hazardous chemicals (fume hoods) and Lot 2 - Mixers*

Contracting authority received a request from an interested economic entity with the following content:

1st REQUEST:

""Group 1 - digesters,
required characteristics 5 and 11 – there is no mention of the regulation of airflow through the a cappella group, nor the possible requirement to be from PPsEL

Is it necessary to put it in the offer, or to stay only on the indication and alarm? The price difference is significant!

Required characteristic 9 - squirt chamber. Does the specified chamber connect to the digester exhaust or does it have to have its own fan??"

THE CONTRACTING AUTHORITY'S RESPONSE:

Digesters must be equipped with a system to control and maintain the amount of air in circulation (airflow control via PPsEL clapboard).

The spray chamber is connected to the digester exhaust.

2nd *REQUEST:*

"According to the open tender "European structural and investment funds", we are interested in offering the project (RINO SPRAY, contract number KK.01.2.1.02.0111).
Given the current situation on the market, we have not been able to obtain all the necessary documentation, that is, all the necessary offers from our equipment suppliers.
Please allow us to submit an offer after the specified deadline for submitting the offer.
The offer would, of course, be submitted by 31.10.2021 if it is acceptable for you."

THE CONTRACTING AUTHORITY'S RESPONSE:

Any Tender received outside the deadline for submission of tenders is incorrect and will not be taken into account.
However, following the request of the interested economic entity, the Contracting authority extends the deadline for submission of tenders until 02.11.2021 until 2:00 PM.

The previous response and 1st modification of the Tender Submission Invitation the Contracting Authority publishes on the same website where the basic documentation was published and it is an integral part of the tender documentation.

The contracting authority invites all interested economic operators to read and follow all the explanations published as part of this procurement procedure.

Procurement Committee